

Supporting information

Characterization of a New Iron Isotopic Reference Material Using Two Independent Methods for Instrumental Mass Bias Correction

Panshu Song,^{a,b} Jiayuan Lou,^c Jianying Zhang,^{a,b} Tianheng Gao,^{a,b} Yuanjing Zhou,^{a,b} and Tongxiang Ren^{a,b,*}

^aDivision of Chemical Metrology and Analytical Chemistry, National Institute of Metrology, Beijing 100029, P. R. China

^bInnovation Center for Advanced Traceability Technology (ICATT), Beijing 10029, P. R. China

^cInstitute of Earth Sciences, China University of Geosciences, Beijing 100083, P. R. China

Table S1. MC-ICPMS operation conditions

Parameters	Values
RF power	1200 W
Ar cooling gas flow rate	16.0 L/min
Aux gas	0.80 L/min
Sample gas	1.05 L/min
Sample cones	Standard sample cone
Skimmer cones	X skimmer cone
Sample uptake rate	0.1 mL/min
Mass resolution	Medium
	L2 ⁵³ Cr
	L1 ⁵⁴ Fe+ ⁵⁴ Cr
	C ⁵⁶ Fe
Cup configuration	H1 ⁵⁷ Fe
	H2 ⁵⁸ Fe+ ⁵⁸ Ni
	H4 ⁶⁰ Ni

Table S2. Mass fraction of impurity elements in the raw material of NIM-RM 2712 determined by GDMS

Element	Mass fraction ($\mu\text{g g}^{-1}$)	Element	Mass fraction ($\mu\text{g g}^{-1}$)	Element	Mass fraction ($\mu\text{g g}^{-1}$)
Li	0.004	Se	0.133	Sm	0.003
Be	0.002	Br	0.272	Eu	<0.001
B	1.814	Rb	0.001	Gd	<0.001
F	0.112	Sr	0.001	Tb	<0.001
Na	2.543	Y	<0.001	Dy	0.001
Mg	0.013	Zr	0.179	Ho	0.010
Al	0.037	Nb	0.042	Er	0.001
Si	0.140	Mo	0.136	Tm	<0.001
P	6.799	Ru	0.012	Yb	<0.001
S	5.118	Rh	<0.001	Lu	<0.001
Cl	2.624	Pd	0.007	Hf	0.002
K	0.087	Ag	0.044	Ta	0.008
Ca	0.042	Cd	0.012	W	0.008
Sc	<0.001 ^a	In	0.009	Re	0.002
Ti	0.004	Sn	0.002	Os	0.005
V	0.029	Sb	0.001	Ir	<0.001
Cr	2.359	Te	0.002	Pt	0.002
Mn	0.713	I	0.025	Au	0.346
Co	0.336	Cs	0.151	Hg	0.003
Ni	0.035	Ba	0.001	Tl	<0.001
Cu	0.039	La	0.001	Pb	0.017
Zn	2.818	Ce	0.017	Bi	0.002
Ga	0.146	Pr	0.003	Th	<0.001
Ge	0.066	Nd	<0.001	U	<0.001
As	0.054				

^a Below the detection limit of the method.





Fig. S1. Stability test results of (a) $\delta^{56}\text{Fe}_{\text{IRMM-014}}$ and (b) $\delta^{57}\text{Fe}_{\text{IRMM-014}}$ in NIM-RM 2712 over 15 months. Each test result (gray square) consists of three single measurements (black cycles) from different bottles. The error bars represented the measurement uncertainty for a single δ measurement. The solid lines represented the certified values of NIM-RM 2712 with the dash lines were their expanded uncertainty of $k=2$.